

ARCS Instrument Replacement Form

System: Anemomter WND1 (low)____ **Replacement**

Fill out the following form when replacing instruments at an ARCS site:

Site: Manus_____

OLD Configuration

Part Number	Serial Number	Location	Configuration Version	Change Item 14	Change Item 15	Date GMT	Time Start GMT	Time End GMT
05106/08234	18426/44680	SMET1	V951025.00	0.09795	0.26	03/01/99		

NEW Configuration

Part Number	Serial Number	Location	Configuration Version	Change Item 14	Change Item 15	Copy Config. Logger to PC	Save to EPROM(E)	Update Version #
05106/08234	31102/58342	SMET1	V990301.00	0.09708	0.36	X	X	X

Replacement completed by: Fred Helsel_____

NOTES:

Old Part remains on site as spare: ☐

Old Part packed for shipment to: AIS_____

Other old part disposition (explain):

Other NOTES:

Replace anemometer for repair (speed dropouts observed)